



ENVIRONMENTAL HEALTH AND SAFETY | HAZARDOUS MATERIALS

Satellite Accumulation Area (SAA) Inspection Log

Chemical waste containers must be closed at all times unless waste is being added. Containers must be compatible with waste, in good condition and not overfilled. Secondary containment is required for incompatible wastes, and waste stored outside a cabinet. Do not exceed 10 gallons per SAA with no more than 1 quart of acutely toxic waste. No more than 2 gallons of flammable waste can be stored outside of a flammables cabinet. Inspect each SAA weekly, and record the results on the SAA Inspection Log. Deficiencies must be noted and immediately corrected. Retain completed SAA Inspection Logs for 3 years. Additional information can be found at Environmental Health and Safety: www.ucdenver.edu/hazmat.

Principal Investigator _____ SAA Number _____

Building name _____ Room number _____ Laboratory phone number _____

Date	Containers leaking?		Containers closed?		Containers labeled?		Chemicals segregated?		Deficiencies and Corrective Actions Taken	Inspector's Initials
	Yes	No	Yes	No	Yes	No	Yes	No		